

calibration certificate



SenseAnywhere BV | Emmerblok 18 | 4751XE Oud Gastel | The Netherlands

Certificate number: **16040712.B02**

Make: **SenseAnywhere**
Airosensor 20-20-21 /00

Serial number: **10:0F:D7:A3:4E**
Ident. nr.: **---**

Test procedure: The instrument is calibrated according to the comparison method (based on ITS-90).

Data used reference:
Type: **P795**
Certificate nr: **Deutsche Akkreditierungsstelle D-K-15219-01-2582**
Certificate no. 1335

Traceability: The calibration is performed with an instrument that can be traced to (inter-) national standards.

Environmental conditions: The ambient temperature was $20^{\circ}\text{C} \pm 5^{\circ}\text{C}$.

Calibration: **6 april 2016**
Recalibrate before: **7 april 2018**

Results

Reference t_{90} [$^{\circ}\text{C}$]	Instrument t [$^{\circ}\text{C}$]	Deviation $t_{90} - t$ [$^{\circ}\text{C}$]	Specification \pm [$^{\circ}\text{C}$]
0,073	0,16	0,087	0,4
19,981	19,98	-0,001	0,4
45,021	45,02	-0,001	0,4

Conclusion: **The instrument does meet the requirements**

A blue ink handwritten signature, appearing to be "C. Boonen", is written over a horizontal line.

Executed by:

C. Boonen

10:0F:D7:A3:4E --- 16040712.B02

This certificate is issued with reservation that SenseAnywhere does not assume any liability and that applicant gives protection for any liability against others.

SenseAnywhere BV
Emmerblok 18
4751 XE Oud Gastel
Netherlands
E info@senseanywhere.nl

CC.. KvK 20159613
VAT nr.NL8212.37.354B01
IBAN NL78 INGB 0005 0734 62
BIC INGBNL2A