



Mechanical Data

Case Style : flat snap canister

Case Dimensions

Diameter : 60.2mm / 2.38"

Thickness : 15.3mm / 0.6"

Hanging tab : extra 12mm / 0.47"

Mounting hole : 6mm / 0.24" diameter

Weight : 26g / 0.92oz approx.

Features

Memory Size : 8k (Non-volatile)
No. of Readings : 7900 approx.
Resolution : 8 bit
Trigger Start : Reed Relay
Delayed Start : Relative / Actual
up to 45 days
Stop Options : When Full
After n Readings
Never (Wrap around)
Reading Types : Actual, Min, Max.
Logging Interval : 1 sec to 10 days
Offload : While stopped or
when logging in
minute multiples
Alarms : Two, fully
Programmable
Functional Range¹ : -40°C → +85°C /
-40°F → +185°F
IP Rating² : IP54 splashproof
Battery Life³ : 1 year Minimum
Offload Speed : Approx 8 Seconds
Offload Method : Close proximity

Notes:

¹ Functional Range describes the range over which the logger will function, **not** the sensor range over which it will record

² The IP54 rating is only valid when the logger is lying flat on its base.

³ Battery replacement is recommended every year. Use only batteries obtained from your Tinytag supplier. Stop the unit logging before replacing the battery.

Sensor Details

Range : -40°C → +85°C /
-40°F → +185°F
Sensor Type : 10k NTC Thermistor
(Encapsulated)
Sensor location : Internally mounted
Response Time : 10 min to 90% typ.
(in air)
Sensor accuracy : ± 0.2°C / ± 0.36°F
From 0°C → 70°C /
32°F → 158°F
Resolution : < 2°, 60° → 85°C
< 1°, 35° → 60°C
< 0.5°, -25° → 35°C
< 1°, -40° → -25°C

Note:

The response time can be greatly improved by opening the container. However, great care must be taken to avoid touching the electronics or allowing moisture to come into contact with it.

Approvals

This equipment complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause any harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

This product is manufactured by Gemini Data Loggers (UK) Ltd to BS EN ISO 9001:2000 (Certificate No. 6134), and is CE approved to EN50081 part 1:1992 and EN50082 part 1 and 2:1992/95 with any standard leads or probes supplied.



UKAS traceable calibration certificates are available on individual units.

Special Notes

When using the units at low temperatures, condensation may form. Before opening the case to change the battery, allow the units to reach room temperature.

Interface Information and Related Products

To use your Tinytag Data Logger you will require:

Transit base station (ACS-3010) or Tinytag Pad base station (ACS-3020), PC with GLM (SWCD-0009) or Easyview (SW-0500)

Further Related Products:

SER-9513 Tinytag Transit Inductive Service Kit containing battery and Instruction sheet.