# Gemini Data Sheet

# Tinytag Ultra Range H Standard Probe TGU-0073

I ssue 6: 27<sup>th</sup> February 2003 E&OE



# **Mechanical Data**

Case Style: IP54 Plastic shell

#### **Case Dimensions**

Height: 72mm / 2.83" Width: 60mm / 2.36" Depth: 33mm / 1.3" Weight: 50g / 1.8 oz.

#### **Features**

Memory Size: 8k (Non-volatile)

No. of Readings: 7900 Resolution: 8 bit

Delayed Start: Relative / Actual

up to 45 days

Stop Options : When Full

After n Readings Never (Wrap around)

Reading Types: Actual, Min, Max. Logging Interval: 1 sec to 10 days Offload: While stopped or

when logging in minute multiples

Alarms: Two, fully

Functional Range : -40°C → +85°C/

-40°F → +185°F

Programmable

IP Rating: IP54 splashproof Battery Life: Up to 5 years

#### Notes:

Battery replacement recommended every 2 years. It may be replaced with Saft 3.6V 1/2AA Lithium cells. Stop the unit logging before replacing the battery.

Functional Range describes the range over which the logger will function, **not** the sensor range over which it will record.

## **Sensor Details**

Range :  $-30^{\circ}C \rightarrow +50^{\circ}C$ 

-22°F → +122°F

Sensor Type: Standard Probe with

10k NTC Thermistor

Resolution: 0.25 °C at 0 °C /

0.45°F at 32°F

Choice of probe depends on the application. See individual data sheets for detailed probe specifications.

#### **Approvals**

This equipment complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause any harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

This product is manufactured by Gemini Data Loggers (UK) Ltd to BS EN ISO 9001:2000 (Certificate No. 6134), and is CE approved to EN50081 part 1:1992 and EN50082 part 1 and 2:1992/95 with any standard leads or probes supplied.





NAMAS traceable calibration certificates are available on individual units.

# **Special Notes**

When using the units at low temperatures, condensation may form. Before opening the case for any reason, allow the units to reach room temperature.

The IP54 rating is only valid when the rubber connector cap and sensor probe are both securely fitted and the unit is oriented with the hanging tab uppermost.

### **Interface Information and Related Products**

#### To use your Tiny Data Logger you will require:

A suitable sensor probe for your application. (Includes PB-4724, PB-4730, PB-4750, PB-4770). See individual data sheets for details. Tinytag interface cable (CAB-0007), PC with GLM for Windows  $^{TM}$  (SW-0009) or Easyview for Windows  $95^{TM}$  (SW-0500).

#### **Further Related Products:**

SER-9520 Tinytag Ultra Service Kit including battery and seal DWG-ULTRA Drawing with detailed dimensions of Tinytag Ultra casing

TGU-0050 Tinytag Ultra Range H with internal sensor

TGU-0020 Tinytag Ultra Range G for Standard probe (-40°C to +125°C / -40°F to +257°F) lower resolution but wider range.

e-mail: sales@geminidataloggers.com