Gemini Data Sheet

#### TIMITA E KITA TOKSOOO Tengerature Rango Orgerature Rango

#### **Mechanical Data**

Case Style : IP-68 Instrument Height : 82mm / 3.3" Width : 60mm / 2.6" Depth : 40mm / 1.6" Mounting holes : 4.2mm / 0.16" diameter 70mm / 2.8" separation

# Features

Memory Size : 8k (Non-volatile) No. of Readings : 7900 Resolution : 8 bit Delayed Start : Relative / Actual up to 45 days
Trigger Start : Magnetic reed switch
Stop Options : When Full After n Readings
Never (Wrap around)
Reading Types : Actual, Min, Max.
Logging Interval : 1 sec to 10 days
Offload : While stopped or
when logging in
minute multiples
Alarms : Two, fully
Programmable
Functional Range : -40 °C → +85 °C /
-40°F → +185°F
IP Rating : IP68
Battery Life : Up to 3 years
Offload Speed : Approx 20 Seconds
Offload Method : Close Proximity

#### Note:

Functional Range describes the range over which the logger will function, **not** the sensor range over which it will record.

## **Special Notes**

Battery replacement recommended every 2 years. We recommend using only replacement batteries from your Tinytag distributor.

When using the units at low temperatures, condensation may form. Before opening the case for any reason, allow the units to reach room temperature.

## Sensor Details

Tinytag eXtra Internal Temperature

Range :	$-40 \degree C \rightarrow +85 \degree C /$
	-40°F → +185°F
Sensor Type :	10k NTC Thermistor
	(Encapsulated)
Sensor location :	Internally mounted
Response Time :	15 min to 90% in air
	(Approx)
Sensor accuracy :	± 0.2°C/ ± 0.36°F
	From $0^{\circ}C \rightarrow 70^{\circ}C/$
	32°F → 158°F
Resolution	: <2°, 60° → 85°C
	<1°, 35° → 60°C
	<0.5°25° → 35°C
	<1°, -40° → -25°C
	,

### Approvals

This equipment complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause any harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

This product is manufactured by Gemini Data Loggers (UK) Ltd to BS EN ISO 9001:2000 (Certificate No. 6134), and is CE approved to EN50081 part 1:1992 and EN50082 part 1 and 2:1992/95 with any standard leads or probes supplied.



UKAS traceable calibration certificates are available on individual units.

# **Interface Information and Related Products**

To use your Tinytag Data Logger you will require: Tinytag Pad base station (ACS-3020), PC with GLM (SWCD-0009) or Easyview (SW-0500)

#### **Further Related Products:**

SER-9560 Tinytag Extra Service Kit including battery, seal and silica gel pack.

Gemini Data Loggers Pty Australia Gemini Data Loggers UK Ltd UK Gemini Data Loggers Inc USA

http://www.geminidataloggers.com

e-mail: sales@geminidataloggers.com

I ssue 9: 27<sup>th</sup> February 2003 E&OE

**TGX-3080**