# Gemini Data Sheet

## Tinyview Plus Range H, Standard Probe TV-0073

Issue 2: 1st July 1999 E&OE



### **Mechanical Data**

Case Style: IP65 noryl Colour : Blue/orange

food grade Diameter: 60mm / 2.36 Length: 84mm / 3.31 Width: 76mm / 2.99

Depth: 35mm / 1.38 Weight: 85g / 3 oz.

#### **Features**

Memory Size: 16k (Non-volatile) Display: 4 digits + indicators

No. of Readings: 14800 (approx) Resolution: 8 bit

Trigger Start: Magnetic reed switch Delayed Start: Relative / Actual

up to 45 days

Stop Options: When Full

After n Readings Never (Wrap around)

0 0 0 0

Reading Types: Actual, Min, Max. Logging Interval: 1 sec to 10 days Offload: While stopped or

when logging in

minute multiples

Alarms: Two, fully Programmable

Functional Range: -40°C to +85°C/ -40°F to +185°F

Display Functions: -20°C to +70°C /

-4°F to 158°F

IP Rating: IP65 splashproof Battery Life: Up to 3 years

#### Notes:

Battery replacement is recommended every 2 years. It may be replaced with Saft 3.7V 1/2AA Lithium cells. Stop the unit logging before replacing the battery. Functional Range describes the range over which the logger will function, not the sensor range over which it will record.

#### Sensor Details

Range: -30°C to +50°C /

-22°F to +122°F

Sensor Type: Standard Probe with 10k NTC Thermistor

Logger resolution: 0.25°C at 0°C/

0.45°F at 32°F

Display resolution: 0.5°C/

Sensor Probe choice depends on the application. See individual data sheets for detailed probe specifications.

### **Approvals**

This equipment complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause any harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

This product is manufactured within the Orion Group to ISO EN 9002 part 2 (Certificate No. 6134), and is CE approved to EN50081 part 1:1992 and EN50082 part 1 and 2:1992/95 with any standard leads or probes supplied.





NAMAS traceable calibration certificates are available on individual units.

# **Special Notes**

When using the units at low temperatures, condensation may form. Before opening the case for any reason, allow the units to reach room temperature. The IP65 rating is only valid when the connector cap is fitted.

Tinyview has the ability to remind the user when calibration is due. The unit can be given a yearly or 6 monthly calibration service, which includes certificate, new battery and seals along with re-setting of the calibration reminder indicator to the next calibration due date.

#### **Interface Information and Related Products**

#### To use your Tiny Data Logger you will require:

A suitable sensor probe for your application. (Includes PB-4724, PB-4730, PB-4750, PB-4770). See individual data sheets for details. Tinytag interface cable (CAB-0007), PC with GLM for Windows ™ (SW-0009) or Easyview for Windows 95™ (SW-0500); or Psion handheld computer, with Tinylink version 2.0 or above (SW-1101) and suitable cable (See Tinylink Data Sheet)

### **Further Related Products:**

ACS-6000 Trigger Start Magnet

SER-9550 Tinyview Service Kit including battery, seal and silica gel pack.

Tinyview Range G (-40°C to +125°C / -40°F to +257°F) for Standard probe. lower resolution but wider range TV-0020

TV-0050

Tinyview Range H  $(-30^{\circ}\text{C to } +50^{\circ}\text{C} / -22^{\circ}\text{F to } +122^{\circ}\text{F})$  with internal sensor Tinyview Range H  $(-30^{\circ}\text{C to } +50^{\circ}\text{C} / -22^{\circ}\text{F to } +122^{\circ}\text{F})$  with stab probe for taking core temperatures TV-0076

### **Applications**

Recording thermometers for chillers and offices

Applications requiring the meeting of Food Directive 92/1, such as E18-150.

Telefon 089 - 800 6940

089 - 800 69429 Fax eMail

mailto:info@bmc.de

